

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061019\
 Data File : VN056186.D
 Acq On : 10 Jun 2019 23:19
 Operator : JC/SP
 Sample : K3267-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP19-CHY08

Quant Time: Jun 11 04:44:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	295875	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	445325	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	431744	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	128867	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	159519	54.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.60%	
35) Dibromofluoromethane	7.59	113	147109	53.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.68%	
50) Toluene-d8	10.09	98	577119	54.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.78%	
62) 4-Bromofluorobenzene	12.40	95	176798	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
Target Compounds						
16) Acetone	3.82	43	4408	3.137	ug/l	94
27) cis-1,2-Dichloroethene	6.83	96	6585	2.105	ug/l	88
40) Benzene	8.03	78	4602	0.396	ug/l	94
44) Trichloroethene	8.83	130	29996	8.976	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061019\
Data File : VN056186.D
Acq On : 10 Jun 2019 23:19
Operator : JC/SP
Sample : K3267-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP19-CHY08

Quant Time: Jun 11 04:44:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

