


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061019\  
 Data File : VN056155.D  
 Acq On : 10 Jun 2019 8:28  
 Operator : JC/SP  
 Sample : BFB  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N060419W.M  
 Title : SW846 8260  
 Last Update : Tue Jun 04 11:02:09 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 18.1      | 4903    | PASS             |
| 75          | 95           | 30           | 60           | 47.3      | 12823   | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 27112   | PASS             |
| 96          | 95           | 5            | 9            | 6.7       | 1809    | PASS             |
| 173         | 174          | 0.00         | 2            | 1.2       | 314     | PASS             |
| 174         | 95           | 50           | 100          | 99.0      | 26850   | PASS             |
| 175         | 174          | 5            | 9            | 6.9       | 1864    | PASS             |
| 176         | 174          | 95           | 101          | 96.1      | 25792   | PASS             |
| 177         | 176          | 5            | 9            | 6.5       | 1683    | PASS             |