

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061021\
 Data File : VN067268.D
 Acq On : 10 Jun 2021 13:33
 Operator : JC/MD
 Sample : VN0610WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0610WBL01

Quant Time: Jun 11 05:38:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061021W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 11 05:27:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	369936	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	667297	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	606321	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	236885	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	331189	47.365	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.740%
35) Dibromofluoromethane	8.027	113	213473	50.202	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.400%
50) Toluene-d8	10.446	98	859941	51.461	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.920%
62) 4-Bromofluorobenzene	12.734	95	338015	48.608	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.220%
Target Compounds						
95) Naphthalene	15.507	128	19333	5.982	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061021\
Data File : VN067268.D
Acq On : 10 Jun 2021 13:33
Operator : JC/MD
Sample : VN0610WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0610WBL01

Quant Time: Jun 11 05:38:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061021W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 11 05:27:41 2021
Response via : Initial Calibration

