

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061022\
 Data File : VN072758.D
 Acq On : 10 Jun 2022 12:12
 Operator : JC\MD
 Sample : VN0610WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0610WBL01

Quant Time: Jun 11 01:52:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	177776	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	333609	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	306414	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	113908	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	257073	53.878	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.760%	
35) Dibromofluoromethane	8.016	113	188487	57.100	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 114.200%	
50) Toluene-d8	10.439	98	608536	46.205	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 92.420%	
62) 4-Bromofluorobenzene	12.727	95	236500	52.330	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 104.660%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061022\
Data File : VN072758.D
Acq On : 10 Jun 2022 12:12
Operator : JC\MD
Sample : VN0610WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0610WBL01

Quant Time: Jun 11 01:52:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 05:17:31 2022
Response via : Initial Calibration

