

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061022\
 Data File : VN072762.D
 Acq On : 10 Jun 2022 14:08
 Operator : JC\MD
 Sample : N3280-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 21ST-CASING-3

Quant Time: Jun 11 01:53:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	166643	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	328846	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	288118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	143529	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	263488	58.912	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 117.820%			
35) Dibromofluoromethane	8.022	113	183985	56.543	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 113.080%			
50) Toluene-d8	10.439	98	620673	47.809	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 95.620%			
62) 4-Bromofluorobenzene	12.727	95	218706	49.094	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 98.180%			
Target Compounds						Qvalue
16) Acetone	4.310	43	4108	1.917	ug/l #	84
95) Naphthalene	15.498	128	24616	2.610	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061022\
Data File : VN072762.D
Acq On : 10 Jun 2022 14:08
Operator : JC\MD
Sample : N3280-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
21ST-CASING-3

Quant Time: Jun 11 01:53:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 05:17:31 2022
Response via : Initial Calibration

