

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061022\
 Data File : VN072766.D
 Acq On : 10 Jun 2022 15:45
 Operator : JC\MD
 Sample : N3293-15 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 40807

Quant Time: Jun 11 01:54:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	171673	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	320284	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	283413	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	113063	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	248200	53.868	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.740%	
35) Dibromofluoromethane	8.022	113	182313	57.527	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 115.060%	
50) Toluene-d8	10.439	98	575643	45.526	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 91.060%	
62) 4-Bromofluorobenzene	12.727	95	219196	50.519	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.040%	
Target Compounds						
					Qvalue	
16) Acetone	4.304	43	11007	4.985	ug/l	# 75
25) 2-Butanone	7.334	43	164136	46.215	ug/l	# 89
40) Benzene	8.457	78	16798	1.258	ug/l	98
52) Toluene	10.504	92	83770	10.051	ug/l	99
67) Ethyl Benzene	11.839	91	27695	1.950	ug/l	100
68) m/p-Xylenes	11.945	106	37089	7.350	ug/l	85
69) o-Xylene	12.274	106	24341	4.798	ug/l	93
70) Styrene	12.292	104	13440	1.686	ug/l	# 63
80) 1,3,5-Trimethylbenzene	13.057	105	9220	0.959	ug/l	97
84) 1,2,4-Trimethylbenzene	13.368	105	44106	4.797	ug/l	93
95) Naphthalene	15.504	128	33174	4.466	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061022\
Data File : VN072766.D
Acq On : 10 Jun 2022 15:45
Operator : JC\MD
Sample : N3293-15 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
40807

Quant Time: Jun 11 01:54:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 05:17:31 2022
Response via : Initial Calibration

