

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061022\
 Data File : VN072770.D
 Acq On : 10 Jun 2022 17:23
 Operator : JC\MD
 Sample : N3280-02RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 21ST-CASING-3RE

Quant Time: Jun 11 01:55:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	163130	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	312483	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	279538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	151924	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	229404	52.396	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.800%	
35) Dibromofluoromethane	8.022	113	166850	53.962	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.920%	
50) Toluene-d8	10.440	98	537501	43.571	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 87.140%	
62) 4-Bromofluorobenzene	12.728	95	211093	49.866	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.740%	
Target Compounds						
						Qvalue
16) Acetone	4.299	43	5287	2.520	ug/l	# 74
95) Naphthalene	15.504	128	20522	2.056	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061022\
Data File : VN072770.D
Acq On : 10 Jun 2022 17:23
Operator : JC\MD
Sample : N3280-02RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
21ST-CASING-3RE

Quant Time: Jun 11 01:55:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 05:17:31 2022
Response via : Initial Calibration

