

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061022\
 Data File : VN072776.D
 Acq On : 10 Jun 2022 19:50
 Operator : JC\MD
 Sample : N3293-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 52322-A

Quant Time: Jun 11 01:56:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 09 05:17:31 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	171635	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	324934	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	285675	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	108330	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	250775	54.439	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.880%	
35) Dibromofluoromethane	8.022	113	183198	56.979	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 113.960%	
50) Toluene-d8	10.439	98	584544	45.568	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 91.140%	
62) 4-Bromofluorobenzene	12.727	95	222744	50.602	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.200%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.369	59	15480	14.087	ug/l	96
16) Acetone	4.293	43	7170	3.248	ug/l	# 79
31) Cyclohexane	8.086	56	13360	2.058	ug/l	# 62
95) Naphthalene	15.498	128	18870	2.651	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061022\
Data File : VN072776.D
Acq On : 10 Jun 2022 19:50
Operator : JC\MD
Sample : N3293-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
52322-A

Quant Time: Jun 11 01:56:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060822W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 09 05:17:31 2022
Response via : Initial Calibration

