

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061024\
 Data File : VN082506.D
 Acq On : 10 Jun 2024 10:46
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 11 01:02:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.721	0.735	-1.9	92	0.00
3 P	Chloromethane	0.817	0.813	0.5	95	0.00
4 C	Vinyl Chloride	0.768	0.790	-2.9#	96	0.00
5 T	Bromomethane	0.441	0.426	3.4	96	0.00
6 T	Chloroethane	0.501	0.506	-1.0	99	0.00
7 T	Trichlorofluoromethane	1.006	1.065	-5.9	98	0.00
8 T	Diethyl Ether	0.358	0.374	-4.5	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.588	0.592	-0.7	96	0.00
10 T	Methyl Iodide	0.591	0.644	-9.0	97	0.00
11 T	Tert butyl alcohol	0.127	0.117	7.9	86	0.00
12 CM	1,1-Dichloroethene	0.546	0.570	-4.4#	96	0.00
13 T	Acrolein	0.094	0.098	-4.3	92	0.00
14 T	Allyl chloride	1.169	1.129	3.4	94	0.00
15 T	Acrylonitrile	0.328	0.336	-2.4	93	0.00
16 T	Acetone	0.324	0.305	5.9	83	0.00
17 T	Carbon Disulfide	1.764	1.749	0.9	97	0.00
18 T	Methyl Acetate	0.807	0.846	-4.8	101	0.00
19 T	Methyl tert-butyl Ether	1.996	2.083	-4.4	95	0.00
20 T	Methylene Chloride	0.653	0.674	-3.2	100	0.00
21 T	trans-1,2-Dichloroethene	0.595	0.626	-5.2	100	0.00
22 T	Diisopropyl ether	2.441	2.520	-3.2	94	0.00
23 T	Vinyl Acetate	1.925	1.975	-2.6	90	0.00
24 P	1,1-Dichloroethane	1.258	1.297	-3.1	98	0.00
25 T	2-Butanone	0.491	0.482	1.8	89	0.00
26 T	2,2-Dichloropropane	1.054	1.110	-5.3	97	0.00
27 T	cis-1,2-Dichloroethene	0.702	0.754	-7.4	100	0.00
28 T	Bromochloromethane	0.586	0.487	16.9	83	0.00
29 T	Tetrahydrofuran	0.315	0.324	-2.9	92	0.00
30 C	Chloroform	1.202	1.295	-7.7#	102	0.00
31 T	Cyclohexane	1.261	1.128	10.5	91	0.00
32 T	1,1,1-Trichloroethane	1.066	1.152	-8.1	99	0.00
33 S	1,2-Dichloroethane-d4	0.790	0.733	7.2	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.308	0.330	-7.1	96	0.00
36 T	1,1-Dichloropropene	0.516	0.566	-9.7	98	0.00
37 T	Ethyl Acetate	0.577	0.610	-5.7	90	0.00
38 T	Carbon Tetrachloride	0.523	0.617	-18.0	104	0.00
39 T	Methylcyclohexane	0.565	0.594	-5.1	90	0.00
40 TM	Benzene	1.538	1.737	-12.9	100	0.00
41 T	Methacrylonitrile	0.340	0.357	-5.0	94	0.00
42 TM	1,2-Dichloroethane	0.592	0.642	-8.4	98	0.00
43 T	Isopropyl Acetate	1.023	1.091	-6.6	93	0.00
44 TM	Trichloroethene	0.349	0.394	-12.9	102	0.00
45 C	1,2-Dichloropropane	0.397	0.442	-11.3#	100	0.00
46 T	Dibromomethane	0.261	0.293	-12.3	100	0.00
47 T	Bromodichloromethane	0.550	0.637	-15.8	101	0.00
48 T	Methyl methacrylate	0.487	0.536	-10.1	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061024\
 Data File : VN082506.D
 Acq On : 10 Jun 2024 10:46
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 11 01:02:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	102	0.00
50 S	Toluene-d8	1.179	1.149	2.5	86	0.00
51 T	4-Methyl-2-Pentanone	0.581	0.637	-9.6	93	0.00
52 CM	Toluene	0.923	1.069	-15.8#	100	0.00
53 T	t-1,3-Dichloropropene	0.569	0.664	-16.7	98	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.696	-15.6	99	0.00
55 T	1,1,2-Trichloroethane	0.335	0.387	-15.5	103	0.00
56 T	Ethyl methacrylate	0.570	0.649	-13.9	93	0.00
57 T	1,3-Dichloropropane	0.627	0.700	-11.6	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.294	0.304	-3.4	86	0.00
59 T	2-Hexanone	0.426	0.470	-10.3	91	0.00
60 T	Dibromochloromethane	0.370	0.468	-26.5#	104	0.00
61 T	1,2-Dibromoethane	0.345	0.390	-13.0	101	0.00
62 S	4-Bromofluorobenzene	0.421	0.435	-3.3	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.371	0.418	-12.7	107	0.00
65 PM	Chlorobenzene	1.105	1.184	-7.1	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.426	-17.0	105	0.00
67 C	Ethyl Benzene	1.976	2.152	-8.9#	98	0.00
68 T	m/p-Xylenes	0.724	0.824	-13.8	101	0.00
69 T	o-Xylene	0.701	0.781	-11.4	99	0.00
70 T	Styrene	1.165	1.360	-16.7	101	0.00
71 P	Bromoform	0.261	0.327	-25.3#	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.972	4.023	-1.3	98	0.00
74 T	N-amyl acetate	2.185	2.013	7.9	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.145	1.116	2.5	99	0.00
76 T	1,2,3-Trichloropropane	1.052	1.054	-0.2	98	0.00
77 T	Bromobenzene	0.947	0.955	-0.8	102	0.00
78 T	n-propylbenzene	4.704	4.854	-3.2	98	0.00
79 T	2-Chlorotoluene	2.954	2.912	1.4	98	0.00
80 T	1,3,5-Trimethylbenzene	3.212	3.369	-4.9	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.450	0.469	-4.2	102	0.00
82 T	4-Chlorotoluene	2.888	2.931	-1.5	99	0.00
83 T	tert-Butylbenzene	2.810	2.811	-0.0	96	0.00
84 T	1,2,4-Trimethylbenzene	3.286	3.448	-4.9	99	0.00
85 T	sec-Butylbenzene	3.895	3.984	-2.3	97	0.00
86 T	p-Isopropyltoluene	3.165	3.366	-6.4	98	0.00
87 T	1,3-Dichlorobenzene	1.715	1.805	-5.2	104	0.00
88 T	1,4-Dichlorobenzene	1.750	1.774	-1.4	103	0.00
89 T	n-Butylbenzene	2.845	2.938	-3.3	95	0.00
90 T	Hexachloroethane	0.586	0.645	-10.1	104	0.00
91 T	1,2-Dichlorobenzene	1.623	1.677	-3.3	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.235	0.237	-0.9	94	0.00
93 T	1,2,4-Trichlorobenzene	0.910	0.938	-3.1	98	0.00
94 T	Hexachlorobutadiene	0.399	0.393	1.5	100	0.00
95 T	Naphthalene	2.852	2.935	-2.9	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061024\
Data File : VN082506.D
Acq On : 10 Jun 2024 10:46
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 11 01:02:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.911	0.927	-1.8	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6