

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061025\
 Data File : VN086924.D
 Acq On : 10 Jun 2025 13:34
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 11 01:54:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	242189	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	448471	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	388518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	184311	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	155596	47.984	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.960%		
35) Dibromofluoromethane	8.177	113	130634	49.152	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.300%		
50) Toluene-d8	10.571	98	538266	51.160	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	102.320%		
62) 4-Bromofluorobenzene	12.847	95	192204	49.170	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.340%		
Target Compounds						
27) cis-1,2-Dichloroethene	7.500	96	9812	2.736	ug/l	76
44) Trichloroethene	9.359	130	33029	10.756	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061025\
Data File : VN086924.D
Acq On : 10 Jun 2025 13:34
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 11 01:54:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

