

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061119\
 Data File : VN056217.D
 Acq On : 11 Jun 2019 12:41
 Operator : JC/SP
 Sample : K3284-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BPS1-TT-MW308I-20190609

Quant Time: Jun 12 07:02:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	303141	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	445948	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	441016	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	143848	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	158242	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
35) Dibromofluoromethane	7.59	113	147319	53.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.68%	
50) Toluene-d8	10.09	98	575205	54.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.26%	
62) 4-Bromofluorobenzene	12.40	95	183934	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
Target Compounds						
16) Acetone	3.82	43	3391	2.356	ug/l	96
95) Naphthalene	15.14	128	4024	4.391	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061119\
Data File : VN056217.D
Acq On : 11 Jun 2019 12:41
Operator : JC/SP
Sample : K3284-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPS1-TT-MW308I-20190609

Quant Time: Jun 12 07:02:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

