

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061119\
 Data File : VN056220.D
 Acq On : 11 Jun 2019 13:57
 Operator : JC/SP
 Sample : K3284-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BPS1-TT-MW306D-20190609

Quant Time: Jun 12 07:07:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	298634	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	444671	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	431064	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	131065	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	158116	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
35) Dibromofluoromethane	7.59	113	143941	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
50) Toluene-d8	10.09	98	567387	54.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.10%	
62) 4-Bromofluorobenzene	12.40	95	176883	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
Target Compounds						
7) Trichlorofluoromethane	3.01	101	1620	0.391	ug/l	99
12) 1,1-Dichloroethene	3.72	96	1537	0.613	ug/l #	77
44) Trichloroethene	8.83	130	13723	4.112	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061119\
Data File : VN056220.D
Acq On : 11 Jun 2019 13:57
Operator : JC/SP
Sample : K3284-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPS1-TT-MW306D-20190609

Quant Time: Jun 12 07:07:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

