

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061119\
 Data File : VN056223.D
 Acq On : 11 Jun 2019 15:12
 Operator : JC/SP
 Sample : K3284-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-DUP06-20190609

Quant Time: Jun 12 07:16:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	308067	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	452594	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	440839	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	129850	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	160148	52.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.72%	
35) Dibromofluoromethane	7.59	113	148302	53.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.82%	
50) Toluene-d8	10.09	98	581081	54.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.76%	
62) 4-Bromofluorobenzene	12.40	95	176320	48.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.04%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061119\
Data File : VN056223.D
Acq On : 11 Jun 2019 15:12
Operator : JC/SP
Sample : K3284-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-DUP06-20190609

Quant Time: Jun 12 07:16:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

