

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061119\
 Data File : VN056226.D
 Acq On : 11 Jun 2019 16:27
 Operator : JC/SP
 Sample : K3284-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BPS1-TT-MW303D-20190610

Quant Time: Jun 12 07:21:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	279224	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	429949	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	417684	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	124883	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	152482	55.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.00%	
35) Dibromofluoromethane	7.59	113	141029	53.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.92%	
50) Toluene-d8	10.09	98	544853	53.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.36%	
62) 4-Bromofluorobenzene	12.40	95	169605	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
Target Compounds						
44) Trichloroethene	8.83	130	4361	1.352	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061119\
Data File : VN056226.D
Acq On : 11 Jun 2019 16:27
Operator : JC/SP
Sample : K3284-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPS1-TT-MW303D-20190610

Quant Time: Jun 12 07:21:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

