

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061119\
 Data File : VN056223.D
 Acq On : 11 Jun 2019 15:12
 Operator : JC/SP
 Sample : K3284-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-DUP06-20190609

Quant Time: Jun 12 07:16:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	308067	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	452594	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	440839	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	129850	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	160148	52.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.72%	
35) Dibromofluoromethane	7.59	113	148302	53.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.82%	
50) Toluene-d8	10.09	98	581081	54.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.76%	
62) 4-Bromofluorobenzene	12.40	95	176320	48.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.04%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061119\
Data File : VN056223.D
Acq On : 11 Jun 2019 15:12
Operator : JC/SP
Sample : K3284-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-DUP06-20190609

Quant Time: Jun 12 07:16:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. 0.00 min

Lab File: VN056223.D

Acq: 11 Jun 2019 15:12

Instrument :

MSVOA_N

ClientSampleId :


BP-DUP06-20190609

Tot Ion:168 Resp: 308067

Ion Ratio Lower Upper

168 100

99 43.9 37.2 55.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.66



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 52.356 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN056223.D


Acq: 11 Jun 2019 15:12

Tgt Ion: 65 Resp: 160148

Ion Ratio Lower Upper

65 100

67 53.5 0.0 109.2



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO

8.03

80000


60000


40000

20000

0

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN056223.D
 Acq: 11 Jun 2019 15:12




Instrument :
 MSVOA_N
 ClientSampleId :
 BP-DUP06-20190609


Tot Ion:114 Resp: 452594
 Ion Ratio Lower Upper
 114 100
 63 18.1 0.0 39.2
 88 15.1 0.0 31.0



#35
 Dibromofluoromethane
 Concen: 53.407 ug/l
 RT: 7.59 min Scan# 1809
 Delta R.T. 0.00 min
 Lab File: VN056223.D
 Acq: 11 Jun 2019 15:12

Tgt Ion:113 Resp: 148302
 Ion Ratio Lower Upper
 113 100
 111 102.7 81.9 122.9
 192 25.5 19.5 29.3









#50
Toluene-d8
Concen: 54.382 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN056223.D
Acq: 11 Jun 2019 15:12




Instrument :
MSVOA_N
ClientSampleId :
BP-DUP06-20190609


Tot Ion: 98 Resp: 581081
Ion Ratio Lower Upper
98 100
100 64.5 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 48.523 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.00 min
Lab File: VN056223.D
Acq: 11 Jun 2019 15:12

Tgt Ion: 95 Resp: 176320
Ion Ratio Lower Upper
95 100
174 99.3 0.0 191.0
176 96.9 0.0 185.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. 0.00 min

Lab File: VN056223.D

Acq: 11 Jun 2019 15:12

Instrument :

MSVOA_N

ClientSampleId :

BP-DUP06-20190609


Tot Ion:117 Resp: 440839

Ion Ratio Lower Upper

117 100

82 51.0 41.8 62.8




119 31.5 25.4 38.0



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VN0

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3598

Delta R.T. 0.00 min

Lab File: VN056223.D

Acq: 11 Jun 2019 15:12


Tgt Ion:152 Resp: 129850

Ion Ratio Lower Upper

152 100

115 51.3 27.0 81.0

150 156.1 0.0 343.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

