

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN061120\
 Data File : VN061770.D
 Acq On : 12 Jun 2020 5:41
 Operator : JC/MD
 Sample : PB129452TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129452TB

Quant Time: Jun 12 08:44:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	301305	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	502993	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	513369	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	224797	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	165307	46.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.10%	
35) Dibromofluoromethane	7.55	113	133817	52.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.88%	
50) Toluene-d8	10.06	98	621023	52.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.04%	
62) 4-Bromofluorobenzene	12.38	95	214579	49.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.48%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.54	94	579	0.20	ug/l #	42
16) Acetone	3.78	43	17499	11.78	ug/l	100
18) Methyl Acetate	4.29	43	2056	0.63	ug/l #	63
20) Methylene Chloride	4.50	84	12748	2.25	ug/l	96
25) 2-Butanone	6.78	43	1046	0.61	ug/l	97
29) Tetrahydrofuran	7.18	42	1179	1.05	ug/l #	40
31) Cyclohexane	7.62	56	5467	1.08	ug/l #	15
36) 1,1-Dichloropropene	7.63	75	20442	4.59	ug/l #	48
38) Carbon Tetrachloride	7.63	117	27716	6.45	ug/l #	16
43) Isopropyl Acetate	8.13	43	60270	9.34	ug/l #	91
45) 1,2-Dichloropropane	9.24	63	1374	0.42	ug/l #	44
48) Methyl methacrylate	9.15	41	2985	1.03	ug/l #	5
49) 1,4-Dioxane	9.23	88	62	1.38	ug/l #	1
51) 4-Methyl-2-Pentanone	9.95	43	8876	2.44	ug/l #	56
54) cis-1,3-Dichloropropene	9.87	75	34397	5.99	ug/l #	62
56) Ethyl methacrylate	10.51	69	1504	0.30	ug/l #	1
57) 1,3-Dichloropropane	10.74	76	6231	1.09	ug/l #	42
59) 2-Hexanone	10.74	43	64262	24.61	ug/l #	51
71) Bromoform	12.37	173	2027	0.66	ug/l #	1
74) N-amyl acetate	12.01	43	1749	0.34	ug/l #	43
76) 1,2,3-Trichloropropane	12.38	75	102658	25.11	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.71	105	4097	0.30	ug/l	97
81) trans-1,4-Dichloro-2-buten	12.38	75	102658	64.22	ug/l #	15
93) 1,2,4-Trichlorobenzene	14.89	180	1180	0.28	ug/l	92
95) Naphthalene	15.11	128	3885	0.31	ug/l #	92
96) 1,2,3-Trichlorobenzene	15.28	180	1175	0.29	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN061120\
Data File : VN061770.D
Acq On : 12 Jun 2020 5:41
Operator : JC/MD
Sample : PB129452TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129452TB

Quant Time: Jun 12 08:44:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. 0.00 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41

Instrument :

MSVOA_N

ClientSampleId :


PB129452TB

Tgt Ion:168 Resp: 301305

Ion Ratio Lower Upper

168 100

99 51.9 45.0 67.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

7.63



150000

100000

50000

0

Time--> 7.50 7.55 7.60 7.65 7.70



#5

Bromomethane

Concen: 0.20 ug/l

RT: 2.54 min Scan# 239

Delta R.T. 0.01 min

Lab File: VN061770.D


Acq: 12 Jun 2020 5:41

Tgt Ion: 94 Resp: 579

Ion Ratio Lower Upper

94 100

96 37.8 74.8 112.2#



Abundance Ion 93.90 (93.60 to 94.60): VN0

Ion 95.90 (95.60 to 96.60): VN0

2.54

400


300


200

100

0

Time--> 2.50 2.52 2.54 2.56





#16

Acetone

Concen: 11.78 ug/l

RT: 3.78 min Scan# 623

Delta R.T. 0.00 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41

Instrument :

MSVOA_N

ClientSampleId :





PB129452TB

Tgt Ion: 43 Resp: 17499

Ion Ratio Lower Upper

43 100

58 33.9 27.2 40.8



#18

Methyl Acetate

Concen: 0.63 ug/l

RT: 4.29 min Scan# 782

Delta R.T. 0.01 min

Lab File: VN061770.D




Acq: 12 Jun 2020 5:41


Tgt Ion: 43 Resp: 2056

Ion Ratio Lower Upper

43 100

74 9.3 23.2 34.8#









#20
Methylene Chloride
Concen: 2.25 ug/l
RT: 4.50 min Scan# 849
Delta R.T. 0.00 min
Lab File: VN061770.D
Acq: 12 Jun 2020 5:41

Instrument :
MSVOA_N
ClientSampled :
PB129452TB

Tgt Ion: 84 Resp: 12748
Ion Ratio Lower Upper
84 100
49 102.5 85.8 128.8
51 28.6 26.2 39.2
86 63.7 52.6 78.8




Abundance Ion 83.90 (83.60 to 84.60): VNO
Ion 48.90 (48.60 to 49.60): VNO
Ion 50.90 (50.60 to 51.60): VNO
Ion 85.90 (85.60 to 86.60): VNO






#25
2-Butanone
Concen: 0.61 ug/l
RT: 6.78 min Scan# 1557
Delta R.T. -0.01 min
Lab File: VN061770.D
Acq: 12 Jun 2020 5:41

Tgt Ion: 43 Resp: 1046
Ion Ratio Lower Upper
43 100
72 33.9 26.0 39.0



Abundance Ion 43.00 (42.70 to 43.70): VNO
Ion 72.00 (71.70 to 72.70): VNO





#29

Tetrahydrofuran

Concen: 1.05 ug/l

RT: 7.18 min Scan# 1680

Delta R.T. 0.01 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41

Instrument :

MSVOA_N

ClientSampleId :

PB129452TB





Tgt Ion: 42 Resp: 1179

Ion Ratio Lower Upper

42 100

72 0.0 43.6 65.4#

71 21.4 41.2 61.8#



#31

Cyclohexane

Concen: 1.08 ug/l

RT: 7.62 min Scan# 1819

Delta R.T. 0.01 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41




Tgt Ion: 56 Resp: 5467


Ion Ratio Lower Upper

56 100

69 172.5 27.8 41.8#

84 132.0 79.9 119.9#









#33
 1,2-Dichloroethane-d4
 Concen: 46.05 ug/l
 RT: 7.99 min Scan# 1933
 Delta R.T. 0.00 min
 Lab File: VN061770.D
 Acq: 12 Jun 2020 5:41




Instrument :
 MSVOA_N
 ClientSampleId :
 PB129452TB


Tgt Ion: 65 Resp: 165307
 Ion Ratio Lower Upper
 65 100
 67 53.7 0.0 104.8



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.55 min Scan# 2108
 Delta R.T. -0.00 min
 Lab File: VN061770.D
 Acq: 12 Jun 2020 5:41

Tgt Ion: 114 Resp: 502993
 Ion Ratio Lower Upper
 114 100
 63 17.8 0.0 35.8
 88 14.6 0.0 31.0









#35
 Dibromofluoromethane
 Concen: 52.44 ug/l
 RT: 7.55 min Scan# 1796
 Delta R.T. 0.00 min
 Lab File: VN061770.D
 Acq: 12 Jun 2020 5:41

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129452TB

Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.9	83.4	125.0
192	20.0	15.4	23.2




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#36
 1,1-Dichloropropene
 Concen: 4.59 ug/l
 RT: 7.63 min Scan# 1820
 Delta R.T. -0.13 min
 Lab File: VN061770.D
 Acq: 12 Jun 2020 5:41

Tgt Ion	Ratio	Lower	Upper
75	100		
110	9.2	19.5	58.5#
77	0.0	24.7	37.1#



Abundance Ion 74.90 (74.60 to 75.60): V
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): V





#38

Carbon Tetrachloride

Concen: 6.45 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. -0.11 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41

Instrument :

MSVOA_N

ClientSampleId :

PB129452TB


Tgt Ion:117 Resp: 27716

Ion Ratio Lower Upper

117 100

119 5.1 76.5 114.7#




121 0.0 24.1 36.1#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 9.34 ug/l

RT: 8.13 min Scan# 1976

Delta R.T. 0.00 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41


Tgt Ion: 43 Resp: 60270

Ion Ratio Lower Upper

43 100

61 23.5 23.6 35.4#



87 17.9 13.0 19.6




Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0





#45

1,2-Dichloropropane

Concen: 0.42 ug/l

RT: 9.24 min Scan# 2321

Delta R.T. 0.15 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41

Instrument :

MSVOA_N

ClientSampleId :


PB129452TB

Tgt Ion: 63 Resp: 1374

Ion Ratio Lower Upper

63 100

65 0.0 24.7 37.1#



Abundance Ion 62.90 (62.60 to 63.60): VN0

Ion 64.90 (64.60 to 65.60): VN0

9.24

800



600

400

200

0

Time-->



#48

Methyl methacrylate

Concen: 1.03 ug/l

RT: 9.15 min Scan# 2293

Delta R.T. -0.02 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41


Tgt Ion: 41 Resp: 2985

Ion Ratio Lower Upper

41 100

69 0.0 83.9 125.9#

39 0.0 52.1 78.1#



Abundance Ion 41.00 (40.70 to 41.70): VN0

Ion 69.00 (68.70 to 69.70): VN0

Ion 39.00 (38.70 to 39.70): VN0

40000


30000


20000

10000

0

Time-->





#49

1,4-Dioxane

Concen: 1.38 ug/l

RT: 9.23 min Scan# 2318

Delta R.T. 0.06 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41

Instrument :

MSVOA_N

ClientSampleId :

PB129452TB


Tgt Ion: 88 Resp: 62

Ion Ratio Lower Upper

88 100

43 1592893.5 23.1 34.7#




58 8058.1 48.3 72.5#



Abundance Ion 88.00 (87.70 to 88.70): VN061770.D (-2206) (-)

Ion 43.00 (42.70 to 43.70): VN061770.D (-2206) (-)

Ion 58.00 (57.70 to 58.70): VN061770.D (-2206) (-)



#50

Toluene-d8

Concen: 52.52 ug/l

RT: 10.06 min Scan# 2577

Delta R.T. 0.00 min

Lab File: VN061770.D


Acq: 12 Jun 2020 5:41

Tgt Ion: 98 Resp: 621023

Ion Ratio Lower Upper



98 100


100 65.5 52.8 79.2



Abundance Ion 98.00 (97.70 to 98.70): VN061770.D (-2484) (-)

Ion 100.00 (99.70 to 100.70): VN061770.D (-2484) (-)









#51
 4-Methyl-2-Pentanone
 Concen: 2.44 ug/l
 RT: 9.95 min Scan# 2544
 Delta R.T. 0.00 min
 Lab File: VN061770.D
 Acq: 12 Jun 2020 5:41




Instrument :
 MSVOA_N
 ClientSampled :
 PB129452TB


Tgt Ion: 43 Resp: 8876
 Ion Ratio Lower Upper
 43 100
 58 15.7 35.0 52.6#



#54
 cis-1,3-Dichloropropene
 Concen: 5.99 ug/l
 RT: 9.87 min Scan# 2519
 Delta R.T. 0.06 min
 Lab File: VN061770.D
 Acq: 12 Jun 2020 5:41

Tgt Ion: 75 Resp: 34397
 Ion Ratio Lower Upper
 75 100
 77 0.0 25.4 38.2#
 39 27.6 33.8 50.6#





#56

Ethyl methacrylate

Concen: 0.30 ug/l

RT: 10.51 min Scan# 2718

Delta R.T. 0.11 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41

Instrument :

MSVOA_N

ClientSampleId :

PB129452TB





Tgt Ion: 69 Resp: 1504

Ion Ratio Lower Upper

69 100

41 4157.0 43.5 65.3#

39 1948.1 27.0 40.6#



#57

1,3-Dichloropropane

Concen: 1.09 ug/l

RT: 10.74 min Scan# 2788

Delta R.T. 0.06 min

Lab File: VN061770.D




Acq: 12 Jun 2020 5:41


Tgt Ion: 76 Resp: 6231

Ion Ratio Lower Upper

76 100

78 0.0 26.3 39.5#









#59
2-Hexanone
Concen: 24.61 ug/l
RT: 10.74 min Scan# 2788
Delta R.T. 0.01 min
Lab File: VN061770.D
Acq: 12 Jun 2020 5:41




Instrument :
MSVOA_N
ClientSampled :
PB129452TB


Tgt Ion: 43 Resp: 64262
Ion Ratio Lower Upper
43 100
58 24.1 30.7 92.1#



#62
4-Bromofluorobenzene
Concen: 49.74 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. 0.00 min
Lab File: VN061770.D
Acq: 12 Jun 2020 5:41

Tgt Ion: 95 Resp: 214579
Ion Ratio Lower Upper
95 100
174 88.9 0.0 172.8
176 86.0 0.0 164.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.38 min Scan# 2988
Delta R.T. 0.00 min
Lab File: VN061770.D
Acq: 12 Jun 2020 5:41




Instrument :
MSVOA_N
ClientSampleId :
PB129452TB


Tgt Ion:117 Resp: 513369
Ion Ratio Lower Upper
117 100
82 50.6 41.5 62.3
119 31.6 25.6 38.4



#71
Bromoform
Concen: 0.66 ug/l
RT: 12.37 min Scan# 3297
Delta R.T. 0.27 min
Lab File: VN061770.D
Acq: 12 Jun 2020 5:41

Tgt Ion:173 Resp: 2027
Ion Ratio Lower Upper
173 100
175 570.4 24.4 73.2#
254 0.0 0.0 0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.32 min Scan# 3591

Delta R.T. 0.00 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41

Instrument :

MSVOA_N

ClientSampled :

PB129452TB


Tgt Ion: 152 Resp: 224797

Ion Ratio Lower Upper

152 100

115 55.7 28.0 84.0




150 156.9 0.0 344.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 0.34 ug/l

RT: 12.01 min Scan# 3183

Delta R.T. -0.04 min

Lab File: VN061770.D

Acq: 12 Jun 2020 5:41

Tgt Ion: 43 Resp: 1749


Ion Ratio Lower Upper

43 100

70 0.0 40.5 60.7#

55 10.7 20.6 31.0#

61 0.0 21.8 32.6#






Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V









#76
 1,2,3-Trichloropropane
 Concen: 25.11 ug/l
 RT: 12.38 min Scan# 3298
 Delta R.T. -0.15 min
 Lab File: VN061770.D
 Acq: 12 Jun 2020 5:41




Instrument :
 MSVOA_N
 ClientSampleId :
 PB129452TB


Tgt Ion: 75 Resp: 102658
 Ion Ratio Lower Upper
 75 100
 77 1.3 0.0 0.0#



#80
 1,3,5-Trimethylbenzene
 Concen: 0.30 ug/l
 RT: 12.71 min Scan# 3401
 Delta R.T. 0.00 min
 Lab File: VN061770.D
 Acq: 12 Jun 2020 5:41

Tgt Ion: 105 Resp: 4097
 Ion Ratio Lower Upper
 105 100
 120 51.8 25.1 75.2









#81
trans-1,4-Dichloro-2-butene
Concen: 64.22 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. 0.10 min
Lab File: VN061770.D
Acq: 12 Jun 2020 5:41




Instrument :
MSVOA_N
ClientSampled :
PB129452TB


Tgt Ion: 75 Resp: 102658
Ion Ratio Lower Upper
75 100
53 0.0 67.8 101.8#
89 0.0 37.1 55.7#



#93
1,2,4-Trichlorobenzene
Concen: 0.28 ug/l
RT: 14.89 min Scan# 4078
Delta R.T. 0.00 min
Lab File: VN061770.D
Acq: 12 Jun 2020 5:41

Tgt Ion: 180 Resp: 1180
Ion Ratio Lower Upper
180 100
182 90.1 47.8 143.3
145 21.6 15.5 46.5









#95
Naphthalene
Concen: 0.31 ug/l
RT: 15.11 min Scan# 4147
Delta R.T. 0.00 min
Lab File: VN061770.D
Acq: 12 Jun 2020 5:41

Instrument :
MSVOA_N
ClientSampleId :
PB129452TB

Tgt Ion:128 Resp: 3885
Ion Ratio Lower Upper
128 100
127 8.8 10.0 15.0#
129 8.5 8.5 12.7#



#96
1,2,3-Trichlorobenzene
Concen: 0.29 ug/l
RT: 15.28 min Scan# 4202
Delta R.T. 0.00 min
Lab File: VN061770.D
Acq: 12 Jun 2020 5:41

Tgt Ion:180 Resp: 1175
Ion Ratio Lower Upper
180 100
182 86.0 47.1 141.3
145 22.3 16.3 48.9

