

Data File : VN061747.D
Acq On : 11 Jun 2020 15:17
Operator : JC/MD
Sample : L2988-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D3-20200608

Quant Time: Jun 22 00:56:41 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	293110	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	489746	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	478071	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	204808	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	160032	45.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.66%	
35) Dibromofluoromethane	7.55	113	136660	55.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.02%	
50) Toluene-d8	10.06	98	589780	51.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.46%	
62) 4-Bromofluorobenzene	12.38	95	200271	47.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.36%	

Target Compounds			Qvalue			
2) Dichlorodifluoromethane	1.83	85	2192	1.781	ug/l	96
9) 1,1,2-Trichlorotrifluoroet	3.72	101	14353	5.120	ug/l	95
27) cis-1,2-Dichloroethene	6.78	96	3927	1.027	ug/l	93
44) Trichloroethene	8.80	130	401089	97.563	ug/l	96
64) Tetrachloroethene	10.60	164	10138	2.253	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN061747.D
Acq On : 11 Jun 2020 15:17
Operator : JC/MD
Sample : L2988-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D3-20200608

Quant Time: Jun 22 00:56:41 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

