

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061120\
 Data File : VN061753.D
 Acq On : 11 Jun 2020 17:37
 Operator : JC/MD
 Sample : L2988-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D4-20200609

Quant Time: Jun 12 08:07:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	281279	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	482700	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	472354	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	211969	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	159085	47.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.94%	
35) Dibromofluoromethane	7.55	113	135452	55.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.64%	
50) Toluene-d8	10.06	98	583530	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
62) 4-Bromofluorobenzene	12.38	95	200389	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.72	101	705745	262.334	ug/l	97
12) 1,1-Dichloroethene	3.70	96	4878	1.697	ug/l #	66
27) cis-1,2-Dichloroethene	6.78	96	3378	0.920	ug/l	83
44) Trichloroethene	8.80	130	93081	22.972	ug/l	92
64) Tetrachloroethene	10.60	164	39523	8.889	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061120\
Data File : VN061753.D
Acq On : 11 Jun 2020 17:37
Operator : JC/MD
Sample : L2988-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D4-20200609

Quant Time: Jun 12 08:07:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

