

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061120\
 Data File : VN061772.D
 Acq On : 12 Jun 2020 6:27
 Operator : JC/MD
 Sample : PB129452ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129452ZHE#02

Quant Time: Jun 12 09:00:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	292319	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	498849	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	513374	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	223276	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	163176	46.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.70%	
35) Dibromofluoromethane	7.55	113	133990	52.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.90%	
50) Toluene-d8	10.06	98	618702	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
62) 4-Bromofluorobenzene	12.38	95	216184	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
Target Compounds						
16) Acetone	3.78	43	17429	12.261	ug/l	94
20) Methylene Chloride	4.50	84	12645	2.336	ug/l	91
43) Isopropyl Acetate	8.13	43	60695	9.485	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061120\
Data File : VN061772.D
Acq On : 12 Jun 2020 6:27
Operator : JC/MD
Sample : PB129452ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129452ZHE#02

Quant Time: Jun 12 09:00:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

