

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061120\
 Data File : VN061742.D
 Acq On : 11 Jun 2020 12:16
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 05:17:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.523	0.533	-1.9	89	0.00
3 P	Chloromethane	0.610	0.508	16.7	78	0.00
4 C	Vinyl Chloride	0.836	0.709	15.2#	79	0.00
5 T	Bromomethane	0.480	0.448	6.7	86	-0.01
6 T	Chloroethane	0.538	0.478	11.2	85	0.00
7 T	Trichlorofluoromethane	1.078	1.010	6.3	89	0.00
8 T	Diethyl Ether	0.330	0.294	10.9	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.478	0.479	-0.2	99	0.00
10 T	Methyl Iodide	0.664	0.669	-0.8	96	0.00
11 T	Tert butyl alcohol	0.079	0.071	10.1	86	0.00
12 CM	1,1-Dichloroethene	0.511	0.465	9.0#	92	0.00
13 T	Acrolein	0.030	0.019	36.7#	81	0.00
14 T	Allyl chloride	0.668	0.559	16.3	82	0.00
15 T	Acrylonitrile	0.221	0.193	12.7	83	0.00
16 T	Acetone	0.211	0.180	14.7	103	0.00
17 T	Carbon Disulfide	1.124	0.923	17.9	82	0.00
18 T	Methyl Acetate	0.539	0.474	12.1	86	0.00
19 T	Methyl tert-butyl Ether	1.741	1.542	11.4	87	0.00
20 T	Methylene Chloride	0.730	0.536	26.6#	93	0.00
21 T	trans-1,2-Dichloroethene	0.576	0.526	8.7	92	0.00
22 T	Diisopropyl ether	1.472	1.324	10.1	86	0.00
23 T	Vinyl Acetate	1.173	0.965	17.7	77	0.00
24 P	1,1-Dichloroethane	0.934	0.849	9.1	89	0.00
25 T	2-Butanone	0.282	0.241	14.5	83	0.00
26 T	2,2-Dichloropropane	0.854	0.812	4.9	97	0.00
27 T	cis-1,2-Dichloroethene	0.652	0.626	4.0	92	0.00
28 T	Bromochloromethane	0.403	0.334	17.1	89	0.00
29 T	Tetrahydrofuran	0.186	0.148	20.4	76	0.00
30 C	Chloroform	1.033	0.967	6.4#	91	0.00
31 T	Cyclohexane	0.840	0.695	17.3	81	0.00
32 T	1,1,1-Trichloroethane	0.941	0.872	7.3	90	0.00
33 S	1,2-Dichloroethane-d4	0.596	0.503	15.6	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.254	0.260	-2.4	86	0.00
36 T	1,1-Dichloropropene	0.443	0.448	-1.1	88	0.00
37 T	Ethyl Acetate	0.357	0.318	10.9	74	0.00
38 T	Carbon Tetrachloride	0.427	0.474	-11.0	94	0.00
39 T	Methylcyclohexane	0.546	0.531	2.7	84	0.00
40 TM	Benzene	1.366	1.403	-2.7	91	0.00
41 T	Methacrylonitrile	0.163	0.137	16.0	81	0.00
42 TM	1,2-Dichloroethane	0.457	0.446	2.4	86	0.00
43 T	Isopropyl Acetate	0.641	0.571	10.9	78	0.00
44 TM	Trichloroethene	0.420	0.439	-4.5	89	0.00
45 C	1,2-Dichloropropane	0.324	0.339	-4.6#	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061120\
 Data File : VN061742.D
 Acq On : 11 Jun 2020 12:16
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 05:17:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.249	0.251	-0.8	89	0.00
47 T	Bromodichloromethane	0.493	0.515	-4.5	90	0.00
48 T	Methyl methacrylate	0.289	0.250	13.5	78	0.00
49 T	1,4-Dioxane	0.004	0.005	-25.0	110	0.00
50 S	Toluene-d8	1.175	1.159	1.4	83	0.00
51 T	4-Methyl-2-Pentanone	0.361	0.333	7.8	78	0.00
52 CM	Toluene	0.891	0.947	-6.3#	92	0.00
53 T	t-1,3-Dichloropropene	0.531	0.551	-3.8	89	0.00
54 T	cis-1,3-Dichloropropene	0.571	0.598	-4.7	91	0.00
55 T	1,1,2-Trichloroethane	0.355	0.371	-4.5	92	0.00
56 T	Ethyl methacrylate	0.496	0.505	-1.8	85	0.00
57 T	1,3-Dichloropropane	0.568	0.590	-3.9	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.212	0.214	-0.9	86	0.00
59 T	2-Hexanone	0.260	0.245	5.8	80	0.00
60 T	Dibromochloromethane	0.404	0.435	-7.7	91	0.00
61 T	1,2-Dibromoethane	0.370	0.390	-5.4	90	0.00
62 S	4-Bromofluorobenzene	0.429	0.417	2.8	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.471	0.512	-8.7	94	0.00
65 PM	Chlorobenzene	1.039	1.111	-6.9	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.427	-10.3	95	0.00
67 C	Ethyl Benzene	1.796	1.924	-7.1#	93	0.00
68 T	m/p-Xylenes	0.712	0.763	-7.2	92	0.00
69 T	o-Xylene	0.677	0.721	-6.5	91	0.00
70 T	Styrene	1.120	1.242	-10.9	93	0.00
71 P	Bromoform	0.298	0.331	-11.1	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.649	3.713	-1.8	93	0.00
74 T	N-amyl acetate	1.149	1.031	10.3	79	0.00
75 P	1,1,2,2-Tetrachloroethane	1.023	0.947	7.4	86	0.00
76 T	1,2,3-Trichloropropane	0.909	0.867	4.6	89	0.00
77 T	Bromobenzene	0.932	0.984	-5.6	95	0.00
78 T	n-propylbenzene	4.074	4.192	-2.9	94	0.00
79 T	2-Chlorotoluene	2.429	2.408	0.9	91	0.00
80 T	1,3,5-Trimethylbenzene	3.050	3.104	-1.8	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.356	0.329	7.6	83	0.00
82 T	4-Chlorotoluene	2.556	2.569	-0.5	92	0.00
83 T	tert-Butylbenzene	2.695	2.680	0.6	91	0.00
84 T	1,2,4-Trimethylbenzene	3.030	3.126	-3.2	92	0.00
85 T	sec-Butylbenzene	3.473	3.507	-1.0	92	0.00
86 T	p-Isopropyltoluene	3.198	3.315	-3.7	93	0.00
87 T	1,3-Dichlorobenzene	1.650	1.765	-7.0	96	0.00
88 T	1,4-Dichlorobenzene	1.680	1.779	-5.9	96	0.00
89 T	n-Butylbenzene	2.679	2.727	-1.8	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061120\
Data File : VN061742.D
Acq On : 11 Jun 2020 12:16
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 12 05:17:37 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.354	0.464	-31.1#	120	0.00
91 T	1,2-Dichlorobenzene	1.576	1.670	-6.0	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.209	0.169	19.1	75	0.00
93 T	1,2,4-Trichlorobenzene	0.922	1.011	-9.7	95	0.00
94 T	Hexachlorobutadiene	0.393	0.440	-12.0	104	0.00
95 T	Naphthalene	2.806	2.748	2.1	82	0.00
96 T	1,2,3-Trichlorobenzene	0.906	0.975	-7.6	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6