


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061120\
 Data File : VN061741.D
 Acq On : 11 Jun 2020 9:49
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Title : SW846 8260
 Last Update : Fri May 29 02:14:10 2020



AutoFind: Scans 3296, 3297, 3298; Background Corrected with Scan 3283

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	9518	PASS
75	95	30	60	50.5	29202	PASS
95	95	100	100	100.0	57797	PASS
96	95	5	9	7.0	4068	PASS
173	174	0.00	2	1.4	681	PASS
174	95	50	100	85.9	49650	PASS
175	174	5	9	7.7	3814	PASS
176	174	95	101	96.1	47738	PASS
177	176	5	9	5.9	2829	PASS