


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061120\
 Data File : VN061764.D
 Acq On : 12 Jun 2020 1:15
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Title : SW846 8260
 Last Update : Fri May 29 02:14:10 2020



AutoFind: Scans 3297, 3298, 3299; Background Corrected with Scan 3284

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	8849	PASS
75	95	30	60	47.7	24242	PASS
95	95	100	100	100.0	50829	PASS
96	95	5	9	7.5	3826	PASS
173	174	0.00	2	1.6	708	PASS
174	95	50	100	87.4	44405	PASS
175	174	5	9	7.8	3464	PASS
176	174	95	101	98.2	43605	PASS
177	176	5	9	6.4	2774	PASS