

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061121\
 Data File : VN067284.D
 Acq On : 11 Jun 2021 13:29
 Operator : JC/MD
 Sample : M2645-01ME
 Misc : 1.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 23892ME

Manual Integrations
 APPROVED

MMDadoda
 6/14/2021 4:10:25 PM

Quant Time: Jun 12 04:25:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061021W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 11 05:27:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	354646	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.965	114	626544	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	583464	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	253954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	311433	46.460	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.920%		
35) Dibromofluoromethane	8.019	113	201373	50.437	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	100.880%		
50) Toluene-d8	10.441	98	802607	51.154	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.300%		
62) 4-Bromofluorobenzene	12.734	95	337490	51.689	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	103.380%		
Target Compounds						
20) Methylene Chloride	5.058	84	18171m	3.743	ug/l	
95) Naphthalene	15.507	128	32073	6.723	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061121\
Data File : VN067284.D
Acq On : 11 Jun 2021 13:29
Operator : JC/MD
Sample : M2645-01ME
Misc : 1.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
23892ME

Manual Integrations
APPROVED

MMDadoda
6/14/2021 4:10:25 PM

Quant Time: Jun 12 04:25:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061021W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 11 05:27:41 2021
Response via : Initial Calibration

