

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061121\
 Data File : VN067308.D
 Acq On : 12 Jun 2021 1:36
 Operator : JC/MD
 Sample : VN0611WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0611WBL02

Quant Time: Jun 12 09:27:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N061121W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 12 09:13:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.657	128	76025	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.968	114	450690	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.747	117	404263	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	242596	30.364	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.200%
60) 4-Bromofluorobenzene	12.734	95	221930	28.435	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.800%
63) Toluene-d8	10.443	98	595285	30.876	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.933%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061121\
Data File : VN067308.D
Acq On : 12 Jun 2021 1:36
Operator : JC/MD
Sample : VN0611WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0611WBL02

Quant Time: Jun 12 09:27:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N061121W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 12 09:13:15 2021
Response via : Initial Calibration

