

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061122\
 Data File : VN072793.D
 Acq On : 11 Jun 2022 17:19
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 13 01:20:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N060622W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jun 06 14:10:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.663	128	30765	30.000	ug/l	0.01
28) 1,4-Difluorobenzene	8.963	114	181001	30.000	ug/l	# 0.00
57) Chlorobenzene-d5	11.745	117	156670	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.434	65	113986	32.487	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 108.300%	
60) 4-Bromofluorobenzene	12.733	95	81426	25.269	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 84.233%	
63) Toluene-d8	10.439	98	268532	30.941	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 103.133%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061122\
Data File : VN072793.D
Acq On : 11 Jun 2022 17:19
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 13 01:20:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N060622W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jun 06 14:10:15 2022
Response via : Initial Calibration

