

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086946.D
 Acq On : 11 Jun 2025 14:22
 Operator : JC\MD
 Sample : Q2251-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VPB182-HYD-20250605

Quant Time: Jun 12 02:11:00 2025
 Quant Title :
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	313297	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	589692	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	504294	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	239780	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	200852	47.882	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.760%
35) Dibromofluoromethane	8.177	113	172903	49.476	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.960%
50) Toluene-d8	10.571	98	683622	49.415	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.820%
62) 4-Bromofluorobenzene	12.847	95	248144	48.278	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.560%
Target Compounds						
60) Dibromochloromethane	11.359	129	4962	1.189	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086946.D
Acq On : 11 Jun 2025 14:22
Operator : JC\MD
Sample : Q2251-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VPB182-HYD-20250605

Quant Time: Jun 12 02:11:00 2025
Quant Title :
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

