

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
 Data File : VN086964.D
 Acq On : 11 Jun 2025 21:01
 Operator : JC\MD
 Sample : Q2263-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW9-MW01S-20250606

Quant Time: Jun 12 01:47:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	212082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	417259	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	370384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	174250	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	152843	53.827	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.660%	
35) Dibromofluoromethane	8.177	113	126870	51.307	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.620%	
50) Toluene-d8	10.571	98	506520	51.743	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.480%	
62) 4-Bromofluorobenzene	12.847	95	181778	49.981	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 99.960%	
Target Compounds						
95) Naphthalene	15.641	128	8833	0.675	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061125\
Data File : VN086964.D
Acq On : 11 Jun 2025 21:01
Operator : JC\MD
Sample : Q2263-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW9-MW01S-20250606

Quant Time: Jun 12 01:47:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

