

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061219\
 Data File : VN056247.D
 Acq On : 12 Jun 2019 14:34
 Operator : JC/SP
 Sample : K3315-06 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP04-20190607

Quant Time: Jun 13 05:12:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	302099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	455653	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	441280	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142045	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	151350	50.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.92%	
35) Dibromofluoromethane	7.59	113	145131	51.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.82%	
50) Toluene-d8	10.09	98	547804	50.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.84%	
62) 4-Bromofluorobenzene	12.40	95	174023	47.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061219\
Data File : VN056247.D
Acq On : 12 Jun 2019 14:34
Operator : JC/SP
Sample : K3315-06 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP04-20190607

Quant Time: Jun 13 05:12:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

