

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061219\
 Data File : VN056266.D
 Acq On : 12 Jun 2019 22:11
 Operator : JC/SP
 Sample : K3315-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE104D2-20190607

Quant Time: Jun 13 06:03:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	352127	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	518872	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	508409	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	169510	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	170771	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
35) Dibromofluoromethane	7.59	113	164924	51.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.62%	
50) Toluene-d8	10.09	98	646274	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
62) 4-Bromofluorobenzene	12.40	95	203420	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
Target Compounds						
16) Acetone	3.82	43	3190	1.908	ug/l	96
27) cis-1,2-Dichloroethene	6.83	96	19360	5.200	ug/l	88
30) Chloroform	7.37	83	3143	0.561	ug/l	93
44) Trichloroethene	8.83	130	119713	30.745	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061219\
Data File : VN056266.D
Acq On : 12 Jun 2019 22:11
Operator : JC/SP
Sample : K3315-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE104D2-20190607

Quant Time: Jun 13 06:03:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

