

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061219\
 Data File : VN056235.D
 Acq On : 12 Jun 2019 8:56
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jun 13 04:44:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jun 06 04:02:53 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	89	0.00
2 M	Dichlorodifluoromethane	1.687	1.848	-9.5	92	0.00
3 M	Chloromethane	1.971	2.007	-1.8	90	0.00
4 M	Vinyl Chloride	2.010	1.998	0.6	86	0.00
5 M	Bromomethane	1.263	1.321	-4.6	88	0.00
6 M	Chloroethane	1.137	1.119	1.6	83	0.00
7 M	Trichlorofluoromethane	2.487	2.629	-5.7	92	0.00
8 T	Diethyl Ether	1.055	0.979	7.2	81	0.00
9	1,1,2-Trichlorotrifluoroeth	1.535	1.680	-9.4	96	0.00
10 M	1,1-Dichloroethene	1.566	1.567	-0.1	88	0.00
11	Methyl Iodide	2.471	2.314	6.4	85	0.00
12	Methyl Acetate	1.882	1.727	8.2	82	0.00
13 M	Acrolein	0.092	0.082	10.9	83	0.00
14 M	Acrylonitrile	0.839	0.782	6.8	83	0.00
15 M	Acetone	0.275	0.340	-23.6	102	0.00
16 M	Carbon Disulfide	3.726	3.967	-6.5	101	0.00
17	Allyl chloride	2.426	2.394	1.3	90	0.00
18 M	Methylene Chloride	1.813	1.788	1.4	88	0.00
19 M	trans-1,2-Dichloroethene	1.672	1.666	0.4	87	0.00
20 T	Diisopropyl ether	5.027	4.887	2.8	85	0.00
21 M	1,1-Dichloroethane	2.921	2.924	-0.1	89	0.00
22 M	cis-1,2-Dichloroethene	1.951	1.915	1.8	86	0.00
23 M	tert-Butyl Alcohol	0.327	0.254	22.3	71	0.00
24 M	Methyl tert-Butyl Ether	4.991	4.657	6.7	83	0.00
25 M	Chloroform	2.917	3.030	-3.9	92	0.00
26	Cyclohexane	2.710	2.681	1.1	88	0.00
27 s	1,2-Dichloroethane-d4	1.863	1.823	2.1	86	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
29	1,1-Dichloropropene	0.406	0.449	-10.6	91	0.00
30 M	2-Butanone	0.214	0.234	-9.3	88	0.00
31	2,2-Dichloropropane	0.378	0.410	-8.5	89	0.00
32 M	1,1,1-Trichloroethane	0.448	0.499	-11.4	91	0.00
33 M	Carbon Tetrachloride	0.380	0.440	-15.8	97	0.00
34 M	Benzene	1.293	1.383	-7.0	87	0.00
35	Methacrylonitrile	0.180	0.185	-2.8	75	0.00
36 M	1,2-Dichloroethane	0.405	0.444	-9.6	90	0.00
37 M	Trichloroethene	0.362	0.391	-8.0	87	0.00
38	Methylcyclohexane	0.514	0.554	-7.8	91	0.00
39 M	1,2-Dichloropropane	0.331	0.360	-8.8	88	0.00
40	Dibromomethane	0.220	0.244	-10.9	91	0.00
41 M	Bromodichloromethane	0.393	0.453	-15.3	97	0.00
42 M	Vinyl Acetate	0.762	0.782	-2.6	84	0.00
43	Ethyl Acetate	0.409	0.410	-0.2	82	0.00
44	Isopropyl Acetate	0.635	0.607	4.4	81	0.00
45 T	1,4-Dioxane	0.007	0.006#	14.3	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061219\
 Data File : VN056235.D
 Acq On : 12 Jun 2019 8:56
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Jun 13 04:44:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jun 06 04:02:53 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.315	0.310	1.6	86	0.00
47	n-amyl Acetate	0.525	0.487	7.2	80	0.00
48 M	t-1,3-Dichloropropene	0.414	0.415	-0.2	88	0.00
49 T	cis-1,3-Dichloropropene	0.484	0.504	-4.1	88	0.00
50 M	1,1,2-Trichloroethane	0.326	0.351	-7.7	88	0.00
51	Ethyl methacrylate	0.493	0.473	4.1	82	0.00
52	1,3-Dichloropropane	0.529	0.562	-6.2	87	0.00
53 M	Dibromochloromethane	0.319	0.374	-17.2	104	0.00
54 M	1,2-Dibromoethane	0.340	0.351	-3.2	86	0.00
55 M	2-Chloroethyl vinyl ether	0.173	0.149	13.9	76	0.00
56 M	Bromoform	0.231	0.266	-15.2	108	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
58 M	4-Methyl-2-Pentanone	0.454	0.447	1.5	84	0.00
59 M	2-Hexanone	0.325	0.329	-1.2	86	0.00
60 S	4-Bromofluorobenzene	0.445	0.431	3.1	85	0.00
61 M	Tetrachloroethene	0.418	0.462	-10.5	92	0.00
62 M	Toluene	1.620	1.658	-2.3	87	0.00
63 S	Toluene-d8	1.361	1.375	-1.0	84	0.00
64 M	Chlorobenzene	0.982	1.014	-3.3	88	0.00
65	1,1,1,2-Tetrachloroethane	0.354	0.385	-8.8	94	0.00
66 M	Ethyl Benzene	1.725	1.772	-2.7	87	0.00
67 M	m/p-Xylenes	0.660	0.688	-4.2	89	0.00
68 M	o-Xylene	0.646	0.661	-2.3	87	0.00
69 M	Styrene	1.050	1.065	-1.4	89	0.00
70	Isopropylbenzene	1.689	1.775	-5.1	89	0.00
71 M	1,1,2,2-Tetrachloroethane	0.509	0.522	-2.6	86	0.00
72	1,2,3-Trichloropropane	0.441	0.460	-4.3	89	0.00
73	Bromobenzene	0.448	0.470	-4.9	91	0.00
74	n-propylbenzene	1.782	1.850	-3.8	91	0.00
75	2-Chlorotoluene	1.105	1.151	-4.2	90	0.00
76	1,3,5-Trimethylbenzene	1.389	1.465	-5.5	90	0.00
77	t-1,4-Dichloro-2-butene	0.126	0.113	10.3	89	0.00
78	4-Chlorotoluene	1.032	1.042	-1.0	91	0.00
79	tert-butylbenzene	1.231	1.293	-5.0	90	0.00
80	1,2,4-Trimethylbenzene	1.351	1.410	-4.4	90	0.00
81	sec-Butylbenzene	1.557	1.656	-6.4	92	0.00
82	p-Isopropyltoluene	1.408	1.461	-3.8	91	0.00
83 M	1,3-Dichlorobenzene	0.699	0.719	-2.9	93	0.00
84 M	1,4-Dichlorobenzene	0.665	0.656	1.4	93	0.00
85	n-Butylbenzene	1.051	0.979	6.9	89	0.00
86 T	Hexachloroethane	0.219	0.249	-13.7	104	0.00
87 M	1,2-Dichlorobenzene	0.710	0.726	-2.3	91	0.00
88	1,2-Dibromo-3-Chloropropane	0.080	0.073	8.8	88	0.00
89	1,2,4-Trichlorobenzene	0.346	0.267	22.8	84	0.00
90	Hexachlorobutadiene	0.259	0.313	-20.8	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061219\
Data File : VN056235.D
Acq On : 12 Jun 2019 8:56
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC020

Quant Time: Jun 13 04:44:04 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N060619W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jun 06 04:02:53 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	0.943	0.645	31.6#	74	0.00
92	1,2,3-Trichlorobenzene	0.372	0.312	16.1	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0