

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061220\
 Data File : VN061783.D
 Acq On : 12 Jun 2020 12:06
 Operator : JC/MD
 Sample : L2988-21DL 4X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE131D2-20200609DL

Quant Time: Jun 13 05:35:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	268886	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	442524	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	429693	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	186816	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	144046	44.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.92%	
35) Dibromofluoromethane	7.55	113	125043	55.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.40%	
50) Toluene-d8	10.06	98	528725	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	12.38	95	182500	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.72	101	131917	51.295	ug/l	96
12) 1,1-Dichloroethene	3.70	96	1463	0.532	ug/l	87
27) cis-1,2-Dichloroethene	6.78	96	5500	1.568	ug/l	77
44) Trichloroethene	8.80	130	100775	27.129	ug/l	92
64) Tetrachloroethene	10.60	164	10953	2.708	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061220\
Data File : VN061783.D
Acq On : 12 Jun 2020 12:06
Operator : JC/MD
Sample : L2988-21DL 4X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE131D2-20200609DL

Quant Time: Jun 13 05:35:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

