

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061220\
 Data File : VN061784.D
 Acq On : 12 Jun 2020 12:30
 Operator : JC/MD
 Sample : L2988-05DL 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE132D4-20200608DL

Manual Integrations
 APPROVED

MMDadoda
 6/15/2020 11:08:32 AM

Quant Time: Jun 13 05:38:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	301193	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	506074	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	490426	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	212757	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	167424	46.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.32%	
35) Dibromofluoromethane	7.55	113	144227	56.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.36%	
50) Toluene-d8	10.06	98	607994	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
62) 4-Bromofluorobenzene	12.38	95	206642	47.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.22%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.71	101	4308m	1.495	ug/l	
27) cis-1,2-Dichloroethene	6.79	96	1653m	0.421	ug/l	
44) Trichloroethene	8.80	130	177838	41.862	ug/l	95
64) Tetrachloroethene	10.61	164	3276	0.710	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061220\
Data File : VN061784.D
Acq On : 12 Jun 2020 12:30
Operator : JC/MD
Sample : L2988-05DL 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D4-20200608DL

Manual Integrations
APPROVED

MMDadoda
6/15/2020 11:08:32 AM

Quant Time: Jun 13 05:38:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

