

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061220\
 Data File : VN061788.D
 Acq On : 12 Jun 2020 14:03
 Operator : JC/MD
 Sample : L2988-15DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D3-20200608DL

Manual Integrations
 APPROVED

MMDadoda
 6/15/2020 11:08:36 AM

Quant Time: Jun 13 05:49:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	298212	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	505465	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	495107	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	218374	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	166933	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
35) Dibromofluoromethane	7.55	113	145710	56.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.66%	
50) Toluene-d8	10.06	98	606820	51.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.14%	
62) 4-Bromofluorobenzene	12.38	95	205982	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.72	101	176894	62.020	ug/l	96
12) 1,1-Dichloroethene	3.70	96	1517m	0.498	ug/l	
27) cis-1,2-Dichloroethene	6.78	96	1541	0.396	ug/l	91
44) Trichloroethene	8.80	130	67567	15.924	ug/l	95
64) Tetrachloroethene	10.60	164	24305	5.215	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061220\
Data File : VN061788.D
Acq On : 12 Jun 2020 14:03
Operator : JC/MD
Sample : L2988-15DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D3-20200608DL

Manual Integrations
APPROVED

MMDadoda
6/15/2020 11:08:36 AM

Quant Time: Jun 13 05:49:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

