

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061220\
 Data File : VN061789.D
 Acq On : 12 Jun 2020 14:27
 Operator : JC/MD
 Sample : L2988-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE132D1-20200608

Quant Time: Jun 13 05:53:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	278004	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	471852	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	473389	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	207321	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	157515	47.56	ug/l	0.00
Spiked Amount						
			Recovery	=	95.12%	
35) Dibromofluoromethane	7.55	113	140472	58.69	ug/l	0.00
Spiked Amount						
			Recovery	=	117.38%	
50) Toluene-d8	10.06	98	576743	51.99	ug/l	0.00
Spiked Amount						
			Recovery	=	103.98%	
62) 4-Bromofluorobenzene	12.38	95	196686	48.60	ug/l	0.00
Spiked Amount						
			Recovery	=	97.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	3.72	101	14240	5.356	ug/l	96
12) 1,1-Dichloroethene	3.70	96	2910	1.024	ug/l	84
19) Methyl tert-butyl Ether	5.00	73	9048	0.935	ug/l #	88
27) cis-1,2-Dichloroethene	6.79	96	6824	1.881	ug/l	85
30) Chloroform	7.33	83	10514	1.831	ug/l	95
44) Trichloroethene	8.80	130	390498	98.589	ug/l	96
64) Tetrachloroethene	10.60	164	10803	2.424	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061220\
Data File : VN061789.D
Acq On : 12 Jun 2020 14:27
Operator : JC/MD
Sample : L2988-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE132D1-20200608

Quant Time: Jun 13 05:53:07 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

