

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061220\
 Data File : VN061796.D
 Acq On : 12 Jun 2020 17:10
 Operator : JC/MD
 Sample : L2989-02 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE105D2-20200609

Quant Time: Jun 13 06:46:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	276795	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	470075	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	465173	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	199840	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	155455	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
35) Dibromofluoromethane	7.55	113	136524	57.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.50%	
50) Toluene-d8	10.06	98	571842	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
62) 4-Bromofluorobenzene	12.38	95	190631	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.71	101	3114	1.176	ug/l	94
44) Trichloroethene	8.80	130	367567	93.150	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061220\
Data File : VN061796.D
Acq On : 12 Jun 2020 17:10
Operator : JC/MD
Sample : L2989-02 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE105D2-20200609

Quant Time: Jun 13 06:46:47 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

