

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061220\
 Data File : VN061797.D
 Acq On : 12 Jun 2020 17:34
 Operator : JC/MD
 Sample : L2989-08 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE122D2-20200610

Quant Time: Jun 13 06:49:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	274951	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	472478	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	459238	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	201423	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	158973	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
35) Dibromofluoromethane	7.55	113	138222	57.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.34%	
50) Toluene-d8	10.06	98	567931	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
62) 4-Bromofluorobenzene	12.38	95	189017	46.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.30%	
Target Compounds						
44) Trichloroethene	8.80	130	389215	98.134	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061220\
Data File : VN061797.D
Acq On : 12 Jun 2020 17:34
Operator : JC/MD
Sample : L2989-08 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE122D2-20200610

Quant Time: Jun 13 06:49:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

