

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061220\
 Data File : VN061802.D
 Acq On : 12 Jun 2020 19:31
 Operator : JC/MD
 Sample : L2989-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE125D2-20200609

Quant Time: Jun 13 07:06:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	268108	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	450020	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	456523	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	199222	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	151890	47.55	ug/l	0.00
Spiked Amount			Recovery	=	95.10%	
35) Dibromofluoromethane	7.55	113	130924	57.35	ug/l	0.00
Spiked Amount			Recovery	=	114.70%	
50) Toluene-d8	10.06	98	557484	52.70	ug/l	0.00
Spiked Amount			Recovery	=	105.40%	
62) 4-Bromofluorobenzene	12.38	95	186909	48.43	ug/l	0.00
Spiked Amount			Recovery	=	96.86%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.83	85	1547	1.636	ug/l	92
9) 1,1,2-Trichlorotrifluoroet	3.72	101	54260	21.160	ug/l	96
12) 1,1-Dichloroethene	3.70	96	14970	5.464	ug/l	90
27) cis-1,2-Dichloroethene	6.78	96	15330	4.382	ug/l	88
30) Chloroform	7.33	83	2445	0.442	ug/l	99
32) 1,1,1-Trichloroethane	7.52	97	2015	0.399	ug/l #	38
38) Carbon Tetrachloride	7.73	117	1961	0.510	ug/l	89
44) Trichloroethene	8.80	130	1163055	307.880	ug/l	96
49) 1,4-Dioxane	9.17	88	1407	34.968	ug/l #	85
55) 1,1,2-Trichloroethane	10.53	97	1470	0.460	ug/l #	90
64) Tetrachloroethene	10.60	164	18276	4.253	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061220\
Data File : VN061802.D
Acq On : 12 Jun 2020 19:31
Operator : JC/MD
Sample : L2989-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE125D2-20200609

Quant Time: Jun 13 07:06:53 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

