

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061223\
 Data File : VN078168.D
 Acq On : 12 Jun 2023 15:54
 Operator : JC\MD
 Sample : 03117-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-26OA

Quant Time: Jun 13 01:34:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	498812	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	951205	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	870657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	293784	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	349929	52.615	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 105.240%		
35) Dibromofluoromethane	8.177	113	310085	51.558	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.120%		
50) Toluene-d8	10.571	98	1132022	48.824	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 97.640%		
62) 4-Bromofluorobenzene	12.853	95	367611	51.001	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 102.000%		
Target Compounds						
16) Acetone	4.448	43	31398	16.402	ug/l	98
20) Methylene Chloride	5.295	84	11721	1.853	ug/l #	91
37) Ethyl Acetate	7.571	43	65769	9.759	ug/l	96
43) Isopropyl Acetate	8.700	43	510442	39.604	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061223\
Data File : VN078168.D
Acq On : 12 Jun 2023 15:54
Operator : JC\MD
Sample : 03117-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-26OA

Quant Time: Jun 13 01:34:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

