

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061223\
 Data File : VN078171.D
 Acq On : 12 Jun 2023 17:06
 Operator : JC\MD
 Sample : 03140-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3-GA

Quant Time: Jun 13 01:35:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	459523	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	862409	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	791372	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	246256	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	329262	53.741	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.480%		
35) Dibromofluoromethane	8.177	113	289796	53.145	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.300%		
50) Toluene-d8	10.576	98	1039087	49.430	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 98.860%		
62) 4-Bromofluorobenzene	12.859	95	320959	49.113	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 98.220%		
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	31242	17.716	ug/l	96
20) Methylene Chloride	5.283	84	11452	1.965	ug/l	91
37) Ethyl Acetate	7.571	43	64689	10.587	ug/l	97
43) Isopropyl Acetate	8.700	43	339487	29.052	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061223\
Data File : VN078171.D
Acq On : 12 Jun 2023 17:06
Operator : JC\MD
Sample : 03140-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3-GA

Quant Time: Jun 13 01:35:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

