



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN061223\  
 Data File : VN078156.D  
 Acq On : 12 Jun 2023 08:53  
 Operator : JC\MD  
 Sample : BFB  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N060623W.M  
 Title : SW846 8260  
 Last Update : Wed Jun 07 05:05:00 2023



AutoFind: Scans 1852, 1853, 1854; Background Corrected with Scan 1843

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 18.7         | 16913      | PASS                |
| 75             | 95              | 30              | 60              | 54.2         | 49027      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 90429      | PASS                |
| 96             | 95              | 5               | 9               | 6.9          | 6261       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.2          | 781        | PASS                |
| 174            | 95              | 50              | 100             | 72.9         | 65939      | PASS                |
| 175            | 174             | 5               | 9               | 7.2          | 4729       | PASS                |
| 176            | 174             | 95              | 101             | 95.0         | 62653      | PASS                |
| 177            | 176             | 5               | 9               | 6.0          | 3731       | PASS                |

Average of 12.847 to 12.859 min.: VN078156.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z   | abund. | m/z   | abund. |
|-------|--------|-------|--------|-------|--------|-------|--------|
| 35.10 | 89     | 48.90 | 893    | 61.05 | 4988   | 71.20 | 56     |
| 35.90 | 354    | 49.05 | 2707   | 62.00 | 5386   | 71.95 | 431    |
| 37.05 | 4038   | 50.05 | 16913  | 63.00 | 3790   | 73.00 | 1676   |
| 38.00 | 4457   | 51.00 | 5557   | 63.70 | 100    | 74.00 | 49027  |
| 39.05 | 2157   | 51.95 | 345    | 64.05 | 594    | 75.00 | 4782   |
| 42.00 | 61     | 52.90 | 56     | 65.00 | 794    | 76.05 | 729    |
| 43.90 | 476    | 54.00 | 59     | 66.95 | 295    | 77.05 | 311    |
| 45.05 | 802    | 54.95 | 211    | 68.00 | 9524   | 78.00 | 3826   |
| 45.85 | 189    | 55.95 | 1724   | 69.00 | 9745   | 78.90 | 1043   |
| 46.95 | 1535   | 57.05 | 2796   | 69.95 | 530    | 79.90 | 3851   |
| 47.85 | 762    | 59.95 | 926    | 70.20 | 176    | 80.90 |        |

Average of 12.847 to 12.859 min.: VN078156.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|-------|--------|--------|--------|--------|--------|--------|--------|
| 81.85 | 988    | 96.85  | 210    | 116.65 | 431    | 130.95 | 135    |
| 83.10 | 179    | 103.90 | 633    | 116.90 | 211    | 133.10 | 72     |
| 86.95 | 3762   | 104.90 | 97     | 117.95 | 409    | 133.70 | 67     |
| 87.95 | 3943   | 105.60 | 133    | 118.70 | 188    | 134.80 | 256    |
| 90.85 | 242    | 105.85 | 280    | 118.95 | 429    | 137.00 | 53     |
| 91.20 | 99     | 106.80 | 90     | 125.20 | 62     | 139.70 | 57     |
| 91.95 | 2982   | 109.80 | 182    | 127.70 | 71     | 140.10 | 60     |
| 92.95 | 4092   | 110.80 | 129    | 128.00 | 397    | 140.80 | 649    |
| 94.00 | 11532  | 112.80 | 75     | 128.80 | 147    | 141.00 | 678    |
| 95.00 | 90429  | 114.75 | 134    | 129.80 | 369    | 141.90 | 96     |
| 96.05 | 6261   | 115.85 | 395    | 130.20 | 80     | 142.10 | 90     |

Average of 12.847 to 12.859 min.: VN078156.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z    | abund. | m/z    | abund. | m/z | abund. |
|--------|--------|--------|--------|--------|--------|-----|--------|
| 142.90 | 995    | 172.05 | 718    | 264.90 | 218    |     |        |
| 144.00 | 71     | 172.60 | 163    | 281.10 | 717    |     |        |
| 144.95 | 198    | 173.05 | 781    | 282.10 | 535    |     |        |
| 147.90 | 93     | 173.95 | 65939  | 282.70 | 60     |     |        |
| 149.95 | 145    | 174.95 | 4729   | 283.10 | 164    |     |        |
| 155.05 | 156    | 175.95 | 62653  |        |        |     |        |
| 156.70 | 64     | 176.95 | 3731   |        |        |     |        |
| 159.10 | 53     | 177.80 | 66     |        |        |     |        |
| 163.10 | 64     | 190.70 | 57     |        |        |     |        |
| 171.05 | 284    | 192.95 | 313    |        |        |     |        |
| 171.90 | 381    | 206.90 | 228    |        |        |     |        |

Instrument :

MSVOA\_N

ClientSampleId :

BFB