

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061224\
 Data File : VN082537.D
 Acq On : 12 Jun 2024 18:36
 Operator : JC\MD
 Sample : P2838-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-196-GW-340-342

Quant Time: Jun 12 19:04:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	159243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	293075	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	290659	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	118732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	134278	53.402	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.800%	
35) Dibromofluoromethane	8.165	113	97861	54.281	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.560%	
50) Toluene-d8	10.571	98	363382	52.572	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.140%	
62) 4-Bromofluorobenzene	12.853	95	119974	48.630	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 97.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061224\
Data File : VN082537.D
Acq On : 12 Jun 2024 18:36
Operator : JC\MD
Sample : P2838-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-196-GW-340-342

Quant Time: Jun 12 19:04:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

