

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061224\
 Data File : VN082542.D
 Acq On : 12 Jun 2024 20:37
 Operator : JC\MD
 Sample : P2840-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BPOW6-9-20240610

Quant Time: Jun 13 01:33:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168257	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	309801	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	299843	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125604	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	138474	52.120	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.240%	
35) Dibromofluoromethane	8.171	113	101426	53.221	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.440%	
50) Toluene-d8	10.571	98	371924	50.903	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.800%	
62) 4-Bromofluorobenzene	12.853	95	124190	47.621	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 95.240%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	3350	3.074	ug/l #	80
52) Toluene	10.630	92	1657	0.290	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061224\
Data File : VN082542.D
Acq On : 12 Jun 2024 20:37
Operator : JC\MD
Sample : P2840-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-9-20240610

Quant Time: Jun 13 01:33:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

