

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
 Data File : VN086972.D
 Acq On : 12 Jun 2025 13:15
 Operator : JC\MD
 Sample : Q2249-01 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-06-6.5-060525

Quant Time: Jun 13 01:26:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	190645	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	365841	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	324312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	154633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	130111	50.973	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 101.940%		
35) Dibromofluoromethane	8.177	113	109990	50.732	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.460%		
50) Toluene-d8	10.571	98	444898	51.836	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.680%		
62) 4-Bromofluorobenzene	12.847	95	155483	48.760	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 97.520%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.406	101	60827	29.266	ug/l	99
16) Acetone	4.453	43	5173	3.822	ug/l	88
27) cis-1,2-Dichloroethene	7.500	96	10861	3.847	ug/l	76
44) Trichloroethene	9.359	130	715258	285.535	ug/l	100
64) Tetrachloroethene	11.112	164	5215	2.541	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
Data File : VN086972.D
Acq On : 12 Jun 2025 13:15
Operator : JC\MD
Sample : Q2249-01 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-06-6.5-060525

Quant Time: Jun 13 01:26:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

