

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
 Data File : VN086973.D
 Acq On : 12 Jun 2025 13:38
 Operator : JC\MD
 Sample : Q2271-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP12-MHK-WC

Quant Time: Jun 13 01:27:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	308710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	601460	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	542008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	259216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	216403	52.356	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.720%	
35) Dibromofluoromethane	8.177	113	181039	50.791	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.580%	
50) Toluene-d8	10.571	98	740754	52.497	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.000%	
62) 4-Bromofluorobenzene	12.847	95	266883	50.908	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.820%	
Target Compounds						
					Qvalue	
16) Acetone	4.465	43	26976	12.307	ug/l	95
18) Methyl Acetate	5.047	43	27430	4.294	ug/l	99
20) Methylene Chloride	5.294	84	9356	2.280	ug/l	95
37) Ethyl Acetate	7.577	43	18778	2.777	ug/l	98
43) Isopropyl Acetate	8.694	43	11803	1.088	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
Data File : VN086973.D
Acq On : 12 Jun 2025 13:38
Operator : JC\MD
Sample : Q2271-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP12-MHK-WC

Quant Time: Jun 13 01:27:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

