

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
 Data File : VN086974.D
 Acq On : 12 Jun 2025 14:00
 Operator : JC\MD
 Sample : Q2274-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-13-MHP-WC

Quant Time: Jun 13 01:27:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	235101	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	450003	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	408636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	193941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	163940	52.082	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.160%		
35) Dibromofluoromethane	8.177	113	135918	50.966	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.940%		
50) Toluene-d8	10.571	98	557074	52.767	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 105.540%		
62) 4-Bromofluorobenzene	12.847	95	195584	49.864	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 99.720%		
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	27428	16.431	ug/l	96
18) Methyl Acetate	5.053	43	28230	5.803	ug/l	97
20) Methylene Chloride	5.295	84	8764	2.804	ug/l	96
37) Ethyl Acetate	7.571	43	18841	3.724	ug/l	98
43) Isopropyl Acetate	8.700	43	12356	1.522	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
Data File : VN086974.D
Acq On : 12 Jun 2025 14:00
Operator : JC\MD
Sample : Q2274-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-13-MHP-WC

Quant Time: Jun 13 01:27:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

