

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
 Data File : VN086979.D
 Acq On : 12 Jun 2025 15:54
 Operator : JC\MD
 Sample : Q2273-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-6

Quant Time: Jun 13 01:29:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	213622	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	414501	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	376760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	177711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	152946	53.475	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.940%	
35) Dibromofluoromethane	8.177	113	125512	51.095	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.200%	
50) Toluene-d8	10.571	98	511198	52.569	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.140%	
62) 4-Bromofluorobenzene	12.847	95	185252	51.275	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.560%	
Target Compounds						
					Qvalue	
16) Acetone	4.454	43	27808	18.334	ug/l	93
18) Methyl Acetate	5.048	43	28573	6.464	ug/l	98
20) Methylene Chloride	5.306	84	8702	3.064	ug/l	90
37) Ethyl Acetate	7.583	43	19081	4.095	ug/l	99
43) Isopropyl Acetate	8.700	43	12458	1.666	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
Data File : VN086979.D
Acq On : 12 Jun 2025 15:54
Operator : JC\MD
Sample : Q2273-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-6

Quant Time: Jun 13 01:29:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

