

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
 Data File : VN086985.D
 Acq On : 12 Jun 2025 18:10
 Operator : JC\MD
 Sample : Q2268-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB-20250605

Quant Time: Jun 13 01:31:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	274527	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	543078	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	482706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	228705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	194567	52.935	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.860%	
35) Dibromofluoromethane	8.177	113	161926	50.312	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.620%	
50) Toluene-d8	10.571	98	660756	51.861	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.720%	
62) 4-Bromofluorobenzene	12.853	95	233413	49.310	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 98.620%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
Data File : VN086985.D
Acq On : 12 Jun 2025 18:10
Operator : JC\MD
Sample : Q2268-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB-20250605

Quant Time: Jun 13 01:31:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

