

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
 Data File : VN086986.D
 Acq On : 12 Jun 2025 18:33
 Operator : JC\MD
 Sample : Q2268-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4-20250605

Quant Time: Jun 13 01:32:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	290913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	573001	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	515947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	245935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	205650	52.798	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	105.600%	
35) Dibromofluoromethane	8.177	113	170765	50.288	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.580%	
50) Toluene-d8	10.571	98	700800	52.132	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	104.260%	
62) 4-Bromofluorobenzene	12.847	95	251186	50.293	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	100.580%	
Target Compounds						
					Qvalue	
30) Chloroform	7.983	83	6861	1.055	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061225\
Data File : VN086986.D
Acq On : 12 Jun 2025 18:33
Operator : JC\MD
Sample : Q2268-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4-20250605

Quant Time: Jun 13 01:32:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

