

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061319\
 Data File : VN056295.D
 Acq On : 13 Jun 2019 20:41
 Operator : JC/SP
 Sample : K3345-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 14 01:15:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	346858	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	572138	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	555520	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	179060	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	194152	56.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.74%	
35) Dibromofluoromethane	7.59	113	178141	50.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.50%	
50) Toluene-d8	10.09	98	705397	52.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.44%	
62) 4-Bromofluorobenzene	12.40	95	229764	50.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.04%	

Target Compounds


						Qvalue
11) Tert butyl alcohol	4.74	59	4883	8.147	ug/l #	74
13) Acrolein	3.80	56	72	0.417	ug/l #	15
16) Acetone	3.83	43	7440	4.517	ug/l #	84
18) Methyl Acetate	4.34	43	3860	0.622	ug/l #	63
30) Chloroform	7.37	83	1649	0.299	ug/l	96
31) Cyclohexane	7.66	56	8441	1.545	ug/l #	3
36) 1,1-Dichloropropene	7.66	75	22342	4.541	ug/l #	48
38) Carbon Tetrachloride	7.66	117	30212	6.661	ug/l #	15
39) Methylcyclohexane	9.08	83	3395	0.554	ug/l	95
52) Toluene	10.16	92	2650	0.285	ug/l	86
67) Ethyl Benzene	11.51	91	292238	15.201	ug/l	100
68) m/p-Xylenes	11.63	106	100994	13.752	ug/l	99
71) Bromoform	12.40	173	1755	3.260	ug/l #	1
73) Isopropylbenzene	12.25	105	63239	2.664	ug/l	100
76) 1,2,3-Trichloropropane	12.40	75	106846	32.429	ug/l #	100
78) n-propylbenzene	12.59	91	53931	3.560	ug/l	99
79) 2-Chlorotoluene	12.69	91	4008	Below Cal	#	39
80) 1,3,5-Trimethylbenzene	12.73	105	22153	1.840	ug/l	99
81) trans-1,4-Dichloro-2-buten	12.40	75	106846	107.994	ug/l #	15
82) 4-Chlorotoluene	12.73	91	2463	0.278	ug/l #	39
83) tert-Butylbenzene	13.04	119	15544	Below Cal	#	72
84) 1,2,4-Trimethylbenzene	13.04	105	118946	10.418	ug/l	100
85) sec-Butylbenzene	13.17	105	4073	0.304	ug/l	85
86) p-Isopropyltoluene	13.51	119	2706	0.225	ug/l	79
90) Hexachloroethane	13.89	117	1221	0.674	ug/l #	1
95) Naphthalene	15.13	128	20995	6.114	ug/l	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061319\
Data File : VN056295.D
Acq On : 13 Jun 2019 20:41
Operator : JC/SP
Sample : K3345-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Jun 14 01:15:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. 0.00 min

Lab File: VN056295.D

Acq: 13 Jun 2019 20:41


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:168 Resp: 346858

Ion Ratio Lower Upper

168 100

99 49.3 37.2 55.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.66



150000

100000

50000

0

Time-->



#11

Tert butyl alcohol

Concen: 8.147 ug/l

RT: 4.74 min Scan# 924

Delta R.T. -0.06 min

Lab File: VN056295.D


Acq: 13 Jun 2019 20:41

Tgt Ion: 59 Resp: 4883

Ion Ratio Lower Upper

59 100

57 0.0 7.8 11.6#



Abundance Ion 59.00 (58.70 to 59.70): VN

Ion 57.00 (56.70 to 57.70): VN

4.74


1500


1000

500

0

Time-->





#13

Acrolein

Concen: 0.417 ug/l

RT: 3.80 min Scan# 629

Delta R.T. 0.20 min

Lab File: VN056295.D

Acq: 13 Jun 2019 20:41


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 56 Resp: 72

Ion Ratio Lower Upper




56 100

55 0.0 55.4 83.2#



Abundance Ion 56.00 (55.70 to 56.70): VN0

Ion 55.00 (54.70 to 55.70): VN0



#16

Acetone

Concen: 4.517 ug/l

RT: 3.83 min Scan# 641

Delta R.T. 0.02 min

Lab File: VN056295.D


Acq: 13 Jun 2019 20:41

Tgt Ion: 43 Resp: 7440

Ion Ratio Lower Upper



43 100


58 23.9 26.6 39.8#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0









#18
Methyl Acetate
Concen: 0.622 ug/l
RT: 4.34 min Scan# 798
Delta R.T. 0.02 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 3860
Ion Ratio Lower Upper
43 100
74 5.9 19.3 28.9#



#30
Chloroform
Concen: 0.299 ug/l
RT: 7.37 min Scan# 1740
Delta R.T. 0.00 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41

Tgt Ion: 83 Resp: 1649
Ion Ratio Lower Upper
83 100
85 61.4 51.8 77.8





#31

Cyclohexane

Concen: 1.545 ug/l

RT: 7.66 min Scan# 1831

Delta R.T. 0.01 min

Lab File: VN056295.D

Acq: 13 Jun 2019 20:41

Instrument :
MSVOA_N
ClientSampled :





Tot Ion: 56 Resp: 8441

Ion Ratio Lower Upper

56 100

69 144.9 25.6 38.4#

84 135.3 67.4 101.0#



#33

1,2-Dichloroethane-d4

Concen: 56.374 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN056295.D




Acq: 13 Jun 2019 20:41


Tgt Ion: 65 Resp: 194152

Ion Ratio Lower Upper

65 100

67 54.6 0.0 109.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.59 min Scan# 2119
Delta R.T. 0.00 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:114 Resp: 572138
Ion Ratio Lower Upper
114 100
63 19.0 0.0 39.2
88 14.6 0.0 31.0




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VN
Ion 87.90 (87.60 to 88.60): VN






#35
Dibromofluoromethane
Concen: 50.749 ug/l
RT: 7.59 min Scan# 1809
Delta R.T. 0.00 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41

Tgt Ion:113 Resp: 178141
Ion Ratio Lower Upper
113 100
111 102.2 81.9 122.9
192 22.6 19.5 29.3



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.541 ug/l
 RT: 7.66 min Scan# 1831
 Delta R.T. -0.13 min
 Lab File: VN056295.D
 Acq: 13 Jun 2019 20:41




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 75 Resp: 22342
 Ion Ratio Lower Upper
 75 100
 110 8.8 19.0 57.0#
 77 0.0 25.1 37.7#



#38
 Carbon Tetrachloride
 Concen: 6.661 ug/l
 RT: 7.66 min Scan# 1832
 Delta R.T. -0.11 min
 Lab File: VN056295.D
 Acq: 13 Jun 2019 20:41

Tgt Ion: 117 Resp: 30212
 Ion Ratio Lower Upper
 117 100
 119 5.7 79.0 118.6#
 121 0.0 24.7 37.1#









#39
Methvlcvclohexane
Concen: 0.554 ug/l
RT: 9.08 min Scan# 2272
Delta R.T. 0.00 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 83 Resp: 3395
Ion Ratio Lower Upper
83 100
55 69.9 59.3 88.9
98 43.0 36.6 54.8




Abundance Ion 83.00 (82.70 to 83.70): VN0
Ion 55.00 (54.70 to 55.70): VN0
Ion 98.00 (97.70 to 98.70): VN0






#50
Toluene-d8
Concen: 52.223 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41

Tgt Ion: 98 Resp: 705397
Ion Ratio Lower Upper
98 100
100 64.2 51.9 77.9



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0





#52

Toluene

Concen: 0.285 ug/l

RT: 10.16 min Scan# 2607

Delta R.T. 0.00 min

Lab File: VN056295.D

Acq: 13 Jun 2019 20:41

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion: 92 Resp: 2650

Ion Ratio Lower Upper




92 100

91 154.8 139.4 209.2



Abundance Ion 92.00 (91.70 to 92.70): VN0

Ion 91.00 (90.70 to 91.70): VN0



#62

4-Bromofluorobenzene

Concen: 50.019 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.00 min

Lab File: VN056295.D

Acq: 13 Jun 2019 20:41


Tgt Ion: 95 Resp: 229764

Ion Ratio Lower Upper

95 100

174 92.2 0.0 191.0



176 88.9 0.0 185.2




Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:117 Resp: 55520
Ion Ratio Lower Upper
117 100
82 53.1 41.8 62.8
119 31.8 25.4 38.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VN
Ion 118.90 (118.60 to 119.60): V






#67
Ethyl Benzene
Concen: 15.201 ug/l
RT: 11.51 min Scan# 3028
Delta R.T. 0.00 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41

Tgt Ion: 91 Resp: 292238
Ion Ratio Lower Upper
91 100
106 31.3 25.1 37.7



Abundance Ion 91.00 (90.70 to 91.70): VN
Ion 106.00 (105.70 to 106.70): V









#68
m/p-Xvlenes
Concen: 13.752 ug/l
RT: 11.63 min Scan# 3064
Delta R.T. 0.01 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion:106 Resp: 100994
Ion Ratio Lower Upper
106 100
91 200.9 159.3 238.9



#71
Bromoform
Concen: 3.260 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.27 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41

Tgt Ion:173 Resp: 1755
Ion Ratio Lower Upper
173 100
175 770.9 24.3 72.8#
254 0.0 0.0 0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN056295.D

Acq: 13 Jun 2019 20:41

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:152 Resp: 179060

Ion Ratio Lower Upper

152 100

115 56.0 27.0 81.0




150 156.1 0.0 343.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#73

Isopropylbenzene

Concen: 2.664 ug/l

RT: 12.25 min Scan# 3258

Delta R.T. 0.00 min

Lab File: VN056295.D


Acq: 13 Jun 2019 20:41

Tgt Ion:105 Resp: 63239

Ion Ratio Lower Upper



105 100


120 27.3 13.6 40.7



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#76

1,2,3-Trichloropropane

Concen: 32.429 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.15 min

Lab File: VN056295.D

Acq: 13 Jun 2019 20:41

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion: 75 Resp: 106846

Ion Ratio Lower Upper




75 100

77 1.1 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0



#78

n-propylbenzene

Concen: 3.560 ug/l

RT: 12.59 min Scan# 3364

Delta R.T. 0.00 min

Lab File: VN056295.D


Acq: 13 Jun 2019 20:41

Tgt Ion: 91 Resp: 53931

Ion Ratio Lower Upper



91 100


120 24.3 12.0 36.0



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 120.00 (119.70 to 120.70): V









#79
 2-Chlorotoluene
 Concen: Below Cal
 RT: 12.69 min Scan# 3395
 Delta R.T. 0.02 min
 Lab File: VN056295.D
 Acq: 13 Jun 2019 20:41




Instrument :
 MSVOA_N
 ClientSampled :


Tot Ion: 91 Resp: 4008
 Ion Ratio Lower Upper
 91 100
 126 0.0 18.1 54.4#



#80
 1,3,5-Trimethylbenzene
 Concen: 1.840 ug/l
 RT: 12.73 min Scan# 3408
 Delta R.T. 0.00 min
 Lab File: VN056295.D
 Acq: 13 Jun 2019 20:41

Tgt Ion: 105 Resp: 22153
 Ion Ratio Lower Upper
 105 100
 120 48.6 24.7 74.1









#81
trans-1,4-Dichloro-2-butene
Concen: 107.994 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.10 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 75 Resp: 106846
Ion Ratio Lower Upper
75 100
53 0.0 68.1 102.1#
89 0.0 36.9 55.3#



#82
4-Chlorotoluene
Concen: 0.278 ug/l
RT: 12.73 min Scan# 3409
Delta R.T. -0.04 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41

Tgt Ion: 91 Resp: 2463
Ion Ratio Lower Upper
91 100
126 0.0 17.8 53.3#









#83
 tert-Butylbenzene
 Concen: Below Cal
 RT: 13.04 min Scan# 3504
 Delta R.T. 0.05 min
 Lab File: VN056295.D
 Acq: 13 Jun 2019 20:41




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion:119 Resp: 15544
 Ion Ratio Lower Upper
 119 100
 91 75.2 30.9 92.8
 134 0.0 13.3 39.8#



#84
 1,2,4-Trimethylbenzene
 Concen: 10.418 ug/l
 RT: 13.04 min Scan# 3504
 Delta R.T. 0.00 min
 Lab File: VN056295.D
 Acq: 13 Jun 2019 20:41

Tgt Ion:105 Resp: 118946
 Ion Ratio Lower Upper
 105 100
 120 46.0 23.0 69.0





#85

sec-Butylbenzene

Concen: 0.304 ug/l

RT: 13.17 min Scan# 3545

Delta R.T. 0.00 min

Lab File: VN056295.D

Acq: 13 Jun 2019 20:41

Instrument :

MSVOA_N


ClientSampled :

Tot Ion:105 Resp: 4073

Ion Ratio Lower Upper




105 100

134 27.6 10.4 31.2



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 0.225 ug/l

RT: 13.51 min Scan# 3650

Delta R.T. 0.22 min

Lab File: VN056295.D

Acq: 13 Jun 2019 20:41


Tgt Ion:119 Resp: 2706

Ion Ratio Lower Upper

119 100

134 39.2 13.4 40.1



91 13.5 10.9 32.9




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V









#90
Hexachloroethane
Concen: 0.674 ug/l
RT: 13.89 min Scan# 3767
Delta R.T. 0.01 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:117 Resp: 1221
Ion Ratio Lower Upper
117 100
201 0.0 50.0 150.0#




Abundance Ion 116.80 (116.50 to 117.50): V
Ion 200.70 (200.40 to 201.40): V



#95
Naphthalene
Concen: 6.114 ug/l
RT: 15.13 min Scan# 4155
Delta R.T. 0.00 min
Lab File: VN056295.D
Acq: 13 Jun 2019 20:41

Tgt Ion:128 Resp: 20995
Ion Ratio Lower Upper
128 100
127 11.5 10.3 15.5
129 10.3 8.7 13.1



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V

