

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061319\
 Data File : VN056281.D
 Acq On : 13 Jun 2019 15:07
 Operator : JC/SP
 Sample : K3345-15DL 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PR-MW-1_061219DL

Manual Integrations
 APPROVED

MMDadoda

6/14/2019 9:37:18 AM

Quant Time: Jun 14 01:49:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 11:02:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	352258	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	578566	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	590510	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	246795	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	193606	55.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.70%	
35) Dibromofluoromethane	7.59	113	179978	50.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.40%	
50) Toluene-d8	10.09	98	722433	52.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.78%	
62) 4-Bromofluorobenzene	12.40	95	266094	57.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.56%	
Target Compounds						
					Qvalue	
16) Acetone	3.83	43	3216	1.923	ug/l	91
39) Methylcyclohexane	9.08	83	9360	1.510	ug/l	92
52) Toluene	10.16	92	52880	5.624	ug/l	99
67) Ethyl Benzene	11.51	91	1161597	56.842	ug/l	99
68) m/p-Xylenes	11.63	106	723443	92.671	ug/l	100
69) o-Xylene	11.95	106	358568	46.695	ug/l	97
73) Isopropylbenzene	12.25	105	130993	5.282	ug/l	98
78) n-propylbenzene	12.59	91	171579	8.218	ug/l	99
80) 1,3,5-Trimethylbenzene	12.73	105	524084	31.584	ug/l	100
84) 1,2,4-Trimethylbenzene	13.04	105	1501884	95.445	ug/l	99
85) sec-Butylbenzene	13.17	105	23051m	1.247	ug/l	
86) p-Isopropyltoluene	13.29	119	19625m	1.186	ug/l	
89) n-Butylbenzene	13.61	91	13014m	1.062	ug/l	
95) Naphthalene	15.13	128	353843	31.518	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061319\
Data File : VN056281.D
Acq On : 13 Jun 2019 15:07
Operator : JC/SP
Sample : K3345-15DL 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PR-MW-1_061219DL

Manual Integrations
APPROVED

MMDadoda

6/14/2019 9:37:18 AM

Quant Time: Jun 14 01:49:40 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N060419W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 11:02:09 2019
Response via : Initial Calibration

